

To: Distribution

From: N. R. Izenberg

Subject: Linearity for CONTOUR imagers

References: (1) N. R. Izenberg, Master Calibration Record Parts 1-4

(2) M. R. Keller, B. Gotwols, Imagers Report SRO-02M-30

Summary

Linearity is determined for Comet Nucleus Tour (CONTOUR) imagers: CONOUR Remote Imager/Spectrometer (CRISP) and CONTOUR Forward Imager (CFI). Parameters were determined from calibration observations of a uniformly illuminated integrating sphere in the APL optical calibration facility (OCF) with varied exposure time. Linearity was determined to be acceptable for flight.

Introduction

CONTOUR'S CFI and CRISP Imager are the two CCD-camera instruments on the CONTOUR spacecraft. Each imager has a 10-filter filter wheel in between the telescope optics and the CCD. Nine of the filters are "spectral" and one is a "clear" or wide-bandpass filter. Accurate knowledge of CCD linearity is required to determine the useful range dynamic range of the instruments for planning flight observations.

To support these analyses, a series of "flatfield" images were collected during the calibration of the instruments. Broad-spectrum, diffuse light was generated in the APL Optical Calibration Facility (OCF) using the facility's integrating sphere.

Test Conditions

The test setup for linearity determinations utilized the OCF "white" sphere fed Halogen and Xenon lamps. The integrating sphere has integrated attenuators to dim the light incoming to the sphere. Each imager took a series of dark and field-filled images at varied exposure times to determine the CCD response as a function of brightness. In order to determine if imager response is linear, we characterize DN accumulation rate (DN with respect to exposure time) with brightness level. All data were corrected for dark current and frame transfer smear to remove additive component of DNs. The additive component would affect results for shorter exposure times

The Dataset for CRISP was taken during the calibration test CIBMA_WLI. CFI linearity data is taken primarily from the CF0AC_WFR test. Calibration observations are documented in the Master Calibration Record (MCR, Ref. 1).



Results

Average raw DN values from field filling images from the integrating sphere, (smear corrected for the shortest exposures), were determined for different exposure times for each imager CCD. The shortest exposure times were used to normalize the change in signal (DN) with the change in exposure time, compared to the Raw DN in each image. The CCD was considered nonlinear as the (Delta DN/ Delta exposure time) value departed from near 1.0. CFI linearity tests (Figure 1, Table 1) included multiple filters at multiple exposure times. CFI was linear to ~3300 DN raw. CRISP linearity tests had on a single filter determined it to be linear to 2000<DN<2700 raw.



Figure 1. Plot of CFI Linearity



Table 1. CFI linearity data for Fig. 1 by filter and exposure time.

Filter	e	xposure/ms	Mean of image/DN	DN Normalized	Delta DN/delta exp. t
	7	81.92	629.4661	391.7361	
		184.32	1118.9355	881.2055	1
		368.64	2000.2596	1762.5296	1.000316874
		552.96	2879.838	2642.108	0.998335476
		737.28	3750.0154	3512.2854	0.987665192
	4	81.92	493.01534	255.28534	
		184.32	811.83285	574.10285	1
		368.64	1384.5637	1146.8337	0.998012329
		737.28	2528.6301	2290.9001	0.996796638
		958.464	3213.2976	2975.5676	0.994222979
	3	6002.048	3960.3744	3722.6444	
		3002.048	3958.205	3720.475	0.609069572
		1002.048	1918.835	1681.105	1.0034372
		737.28	1474.0452	1236.3152	1.0005278
		368.64	856.55404	618.82404	1
		184.32	547,97133	310.24133	0.999695868
		81.92	376 58863	138 85863	
		81.92	247.03895	9,30895	
		184.32	254 17792	16 44792	
		737.28	304 07427	66.34427	1 045245728
		1002 048	326 21066	88 48066	0 968466589
		3002.040	498 71329	260 98329	0.000400000
		6002.040	757 70026	519 97026	0.000100244
	2	81 02	2/0 3/161	11 61161	I
	2	737.28	326 00/35	88 36/35	1 0083311///
		1002 049	356 32444	119 50443	0.000001144
		2002.040	599 35524	350 62524	0.903021043
		6002.040	026 709/7	600.06847	0.99000000
	1	0002.040	930.79047	099.00047	1
	I	01.92	1152 2027	015.9433	1
		92.10	105.2927	1010 4501	1 000010570
		102.4	1207.1001	1019.4001	1.022310373
		142.00	1409.0004	1221.9334	1.003071002
		143.30	1002.9104	1425.1604	1.000040347
		184.32	2071.8000	1834.0706	1.006832045
		204.8	22/3.0038	2035.9338	0.994116946
		225.28	24/7.1500	2239.4266	1.002142247
		245.76	2083.5478	2445.8178	1.015510818
		266.24	2886.8445	2649.1145	1.000284887
		286.72	3088.8863	2851.1563	0.994110377
		307.2	3291.9441	3054.2141	0.999109422
		327.68	3494.9628	3257.2328	0.998917037
		348.16	3691.9291	3454.1991	0.969137291
		368.64	3871.8437	3634.1137	0.885237464
Dark		81.92	240.59683	2.86683	
		163.84	240.34791	2.61791	
		327.68	240.8948	3.1648	
		327.68	240.91159	3.18159	
		655.36	242.0615	4.3315	
		958.464	243,1755	5,4455	







Figure 2. Plot of CRISP Linearity

mean of image/DN	DN Normalized	Delta DN/delta exp.t
285.24	0.28244101	
648.787	0.642420612	1.006712485
1009.91	1	1.0000009
2099.99	2.079383311	1.006261847
3849.07	3.811300017	0.966173349
4079.95	4.039914448	0.063863538
4079.95	4.039914448	0
4079.95	4.039914448	0
4079.95	4.039914448	0
4079.95	4.039914448	0
	mean of image/DN 285.24 648.787 1009.91 2099.99 3849.07 4079.95 4079.95 4079.95 4079.95 4079.95	mean of image/DN DN Normalized 285.24 0.28244101 648.787 0.642420612 1009.91 1 2099.99 2.079383311 3849.07 3.811300017 4079.95 4.039914448 4079.95 4.039914448 4079.95 4.039914448 4079.95 4.039914448 4079.95 4.039914448 4079.95 4.039914448 4079.95 4.039914448 4079.95 4.039914448

Conclusions

CRISP linearity was somewhat low, but linearity for CRISP and CFI was acceptable. In-flight calibrations were planned to update and supplement the results of the on-ground calibrations.



Noam R. Izenberg

Distribution:

B. L. Gotwols D. C. Humm R.W. McEntire S. L. Murchie J. W. Warren SRP Files Archives